

Conductivity Standards



Available Standards

Any combination of the following values up to three per set: 17, 22, 27, 30, 34, 39, 43, 47, 53, 60 and 100%. IACS, nominally.

Materials

Nonferrous, non-magnetic.

Certified Accuracy

+/- .35% IACS or 1% of value, whichever is less, at 20° C, from 60 to 200 kHz (NIST traceable).